



TEST REPORT

Applicant Name & Address : Uni-Trend Technology (China) Limited.
 No 6, Gong Ye Bei 1st Road, Songshan Lake National High-Tech Industrial
 Development Zone, Dongguan City, Guangdong Province, China

Manufacturing Site : Same as applicant

Sample Description

Product : Mini Clamp Meter
 Model No. : UT210A, UT210B, UT210E, UT210C, UT210D
 Electrical Rating : CAT III 300V, CAT II 600V, class II
 Battery supply: 2 x 1.5V, AAA battery

Test standards : EN 61326-1: 2013
 EN 61326-2-2:2013

Test Result : Pass

Conclusion : The submitted samples complied with the above EMC standards.

Remark : None

*****End of Page*****

Prepared and Checked By:

Approved By:

Daniel He

Helen Ma *Signature*

Daniel He
Project Engineer
Intertek Guangzhou

Helen Ma
Team Leader
Intertek Guangzhou

03 June 2016 *Date*

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program. The test report only allows to be revised within three years from its original issued date unless further standard or the requirement was noticed.

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch
Block E, No.7-2 Guang Dong Software Science Park, Caipin Road, Guangzhou Science City, GETDD Guangzhou, China
Tel / Fax: 86-20-8213 9688/86-20-3205 7538



Report No.: 140113019GZU-001
Amendment 4: 2016-6-03

**TEST REPORT
(with Justification)**

Test result:

It is found that the Mini Clamp Meter, model UT210A, UT210B, UT210E, UT210C, UT210D met the requirements of 2014/30/EU directive and EN 61326-1, EN 61326-2-2 standards.

Report revision reason:

This report is the revision of the previous test report 140113019GZU-001 dated 25 Sep. 2015 and shall be used together with it.

The directive has been updated from 2004/108/EC to 2014/30/EU.

Based on engineering judgement, no additional tests are needed.

*****End of report*****